










Some DNA Genotek products may not be available in all geographic regions.
*The OMNIgene®-GUT Dx (OMD-200) device has been granted FDA De Novo classification for in vitro diagnostic use (DEN200040).
OMNIgene®-GUT (OMR-200), OMNIgene®-ORAL (OMR-110/120), OMNIgene®-VAGINAL, OMNIgene®-SALIVA (OMR-610) and OMNIgene®-SKIN devices are for research use only, not for use in diagnostic procedures.
OMNIgene, OMNImet and DNA Genotek are trademarks of DNA Genotek Inc.
Patent (www.dnagenotek.com/legalnotices)
© 2023 DNA Genotek Inc. is a subsidiary of OraSure Technologies, Inc., all rights reserved.
MK-01802 Issue 4/2023-12

DNAgenotek[®]

Microbiome and metabolomics product details

	Stool OMD-200 (DNA)	Stool OMR-200 (DNA)	Stool OMR-205 (DNA & RNA)	Vaginal OMR-130 (DNA & RNA)	Tongue OMR-120 (DNA & RNA)	Gum & plaque OMR-110 (DNA & RNA)	Skin OMR-140 (DNA)	Saliva OMR-610 (DNA & RNA Host and Microbial)	Stool ME-200 (Metabolites)
									
Tare weight for collection device (weight of tube, cap, chemistry and swab, if applicable)	13.69 g	13.82 g	15.80 g	5.33 g	5.33 g	5.33 g	5.62 g	6.04 g	12.31 g
Volume of chemistry	2 mL	2 mL	4 mL	1 mL	1 mL	1 mL	1 mL	1 mL	2.8 mL
Volume of sample collected	0.4-0.5 mL	0.4-0.5 mL	0.4-0.5 mL	Not applicable	Not applicable	Not applicable	Not applicable	1 mL	0.4-0.5 mL
Mean (median) DNA yield/ extraction ¹	9.89 ± 0.23 µg (8.72 µg)	12 ± 1.49 µg (12.56 µg)	4.40 ± 0.52 µg (2.06 µg)	4.29 ± 1.15 µg (2.63 µg)	0.29 ± 0.035 µg (0.297 µg)	0.374 ± 0.036 µg (0.345 µg)	Collection site dependent	1.22 ± 0.12 µg (0.86 µg)	--
Ambient temperature sample stability ²	30 days	60 days	10 days (RNA) and 30 days (DNA)	30 days	30 days	30 days	30 days	21 days	7 days ³

DNA Genotek products are compatible with a wide range of extraction kits.
Please consult the specific product page or contact technical support (support@dnagenotek.com) for additional sample processing protocols and recommendations.

1 Extraction from a 250 µL aliquot.

2 Storage recommendations for samples collected with OMNIgene[®] products from DNA Genotek[®]. DNA Genotek. PD-PR-01036

3 Storage recommendations for samples collected with OMNImet[®]-GUT. DNA Genotek. PD-PR-00987