

USER INSTRUCTIONS

Read all instructions prior to collection

Collection precautions:

Do NOT eat, drink, smoke or chew gum for 30 minutes before giving your saliva sample.

Do NOT remove the plastic film from the cap.

Intended use: For the collection of human DNA from saliva samples.

Contents: Kit contains stabilizing liquid.

Warnings and precautions:

Wash with water if stabilizing liquid comes in contact with eyes or skin. Do NOT ingest. See MSDS at www.dnagenotek.com.

Storage: 15°C 1 25°C

Summary and explanation of the kit:

Oragene•DISCOVER is a self-collection kit that provides the materials and instructions for collecting and stabilizing saliva specimens.

Label legend:

(Ji Consult package insert Collect saliva by (Use by) REF Catalog number

 \triangle Caution consult instructions for use

Storage instructions 15°C ∦ 25°C Manufacturer Lot number LOT

Procedure:


Most people take between 2 and 5 minutes to deliver a sample following steps 1 to 4.



Start giving your saliva sample.



Spit until the amount of liquid saliva (not bubbles) reaches the level shown in picture #2.



Put the container on a flat surface. Screw the cap onto the container. Make sure that the cap is closed tightly, as shown in picture #3c. If you have difficulty closing the cap completely, turn cap slightly counter-clockwise then try again.



Shake gently for 10 seconds. Recycle outer packaging.

For Research Use Only Not for use in diagnostic procedures



Superior samples Proven performance